

**BED DOCKER/LOCATOR** basic unit shall be manufactured by Hospital Systems, Inc., Pittsburg CA 94565 USA, in accordance with shop drawings and documents. The following is a general specification, and components listed may not be present or required in final product.

The BED DOCKER/LOCATOR is listed by Underwriters Laboratories Inc. (UL) in Canada and the United States of America. This product also complies with the seismic requirements of the State of California's Office of Statewide Health Planning and Development (OSPHD). HSI manufactures all products in accordance with the National Fire Protection Association (NFPA), NFPA-99 (current edition) and the National Electric Code (NEC).

The installation contractor is responsible for compliance with all local, state, and federal codes applicable to the installation of medical gas and electrical systems.

### 1. Submittals and Approvals

Job specific shop drawings shall be produced for each project. Drawings of Record will be produced, and emailed at the time of shipment, on request.

HSI will manufacture equipment as per signed approval drawings and verifications documents, which are provided in the job specific submittal package.

### 2. Basic Construction

Enclosure shall be constructed of heat-formed Kydex high impact plastic to form a single-piece, surface mounted unit. Unit dimensions shall be 50.25" x 22.25" x 4.62". Unit shall be factory assembled to include electrical components as shown on the plans and specifications, and wired in accordance with HSI shop drawings. Device cover plates shall be constructed of clear anodized aluminum. Standard color shall be Polar White with other colors available.

### 3. Components

#### **A. Electrical WIRING**

Wire for standard and critical branch power circuits shall be #12 THHN stranded copper wire, 600 volt, with heat resistant thermo-plastic insulation for hot (black) and neutral (white). Grounds shall be #10 THHN stranded copper wire (green). All ground conductors shall be installed in conduit.

#### **ELECTRICAL RECEPTACLES**

Receptacles shall be the type and quantity as shown on the drawings and specified. Unless otherwise noted 2 pole, 3 wire, rated at 20 amps, 120 Volt, Hospital Grade shall be provided.

Also available: BS, DIN, and others.

- **Duplex receptacles** shall be NEMA style 5-15R or 5-20R, color Ivory for use on normal power circuits, and color Red for use on critical branch power circuits.
- **Simplex receptacles** shall be NEMA style 5-15R or 5-20R, color Ivory for use on normal power circuits, and color Red for use on critical branch power circuits.

100% USA Manufactured

Union Made

12 Year Product Warranty

All specifications are nominal and subject to change without notice. Hospital Systems, Inc. makes no warranty, expressed or implied, except for the express written warranty extended on the sale of its products.

Hospital Systems, Inc.—750 Garcia Ave—Pittsburg, CA 94565  
phone: 925.427.7800 fax: 925.427.0800  
[www.HSIheadwalls.com](http://www.HSIheadwalls.com)

### **3. Components Cont.**

- **Safety receptacles** (if required) shall be duplex type, be NEMA style 5-15R or 5-20R, color Ivory for use on normal power circuits, and color Red for use on critical branch power circuits. Receptacles shall limit proper access to energized contracts and shall accept both 2 wire and 3 wire plugs.

#### **GROUNDING JACKS**

Externally accessible grounding jacks shall be provided in the Axiom wall unit as shown in these shop drawings. The solid brass receptacle shall be enclosed in a non-conductive housing and shall be spring loaded, with a twist-to-lock action and shall be manufactured to the requirements of NEC Article 517, and NFPA 99.

#### **NIGHT LIGHTS**

Night light enclosure shall be of extruded aluminum, Night light shall include 120 or 220 V lamp, assembly, and louvered aluminum cover plate. Night light may be continuous, have integral or remote switching.

#### **B. Provisions**

Provisions with blank aluminum cover plates shall be provided as shown on shop drawings.

### **4. Installation**

HSI shall supply a mounting plate with boxes for building service connections. Mounting plate will be supplied in advance of BED DOCKER/LOCATOR units for pre-installation of electrical services. Hardware is to be furnished by the installer or contractor. The installer or contractor shall install back box, make all primary electrical connections, and mount BED DOCKER/LOCATOR. Hardware to mount the unit to the back box shall be provided by the installing contractor to meet local requirements. Instructions shall be furnished by HSI The installer shall make connection of building services to pre-wired junction box as shown on the electrical drawings and as herein specified. Please refer to the installation instructions for further details.

### **5. Ordering Information**

Please visit our website at [www.HSIheadwalls.com](http://www.HSIheadwalls.com) for the contact information of your local sales representative.

**100% USA Manufactured**

**Union Made**

**12 Year Product Warranty**

All specifications are nominal and subject to change without notice. Hospital Systems, Inc. makes no warranty, expressed or implied, except for the express written warranty extended on the sale of its products.