



Security Axiom™ Patient Care Headwall Specifications

AXIOM SECURITY HEADWALL basic unit shall be manufactured by Hospital Systems, Inc., Pittsburg CA 94565 USA, in accordance with shop drawings and documents. The following is a general specification, and components listed may not be present or required in final product.

The SECURITY HEADWALL is listed by Underwriters Laboratories Inc. (UL) in Canada and the United States of America. This product also complies with the seismic requirements of, and is approved by, the State of California's Office of Statewide Health Planning and Development (OSPHD). HSI manufactures all products in accordance with the National Fire Protection Association (NFPA), NFPA-99 (current edition) and National Electric Code (NEC).

The installation contractor is responsible for compliance with all local, state, and federal codes applicable to the installation of medical gas and electrical systems.

1. Submittals and Approvals

Job specific shop drawings shall be produced for each project. These shop drawings will clearly indicate the area of Medical Gas termination, and electrical connection points inside or outside the SECURITY HEADWALL. Drawings of Record will be produced, and emailed at the time of shipment, upon request.

HSI will manufacture equipment as per signed approval drawings and verification documents, which are provided in the job specific submittal package.

2. Basic Construction

Enclosure shall be constructed of powder coated steel or stainless steel as specified, sections to provide a surface or recessed mounted unit. Unit shall be factory assembled to include mechanical and electrical components as shown on the plans and specifications, and wired in accordance with HSI shop drawings. Device cover plates shall be constructed of stainless steel. All secondary wiring within the SECURITY HEADWALL shall be enclosed in conduit and terminate above the ceiling line. All wiring shall be marked, color-coded and tie wrapped to facilitate easy circuit identification.

The security door shall be made of 16ga steel, powder coated or stainless, with a full length hinge.

3. Components

A. Electrical

INTERNAL GROUND BUS

A ground bus shall be included.

100% USA Manufactured

Union Made

12 Year Product Warranty

All specifications are nominal and subject to change without notice. HSI makes no warranty, expressed or implied, except for the express written warranty extended on the sale of its products.

HSI—750 Garcia Ave—Pittsburg, CA 94565
phone: 925.427.7800 fax: 925.427.0800
www.HSIheadwalls.com

3. Components Cont.

ELECTRICAL RECEPTACLES

Receptacles shall be the type and quantity as shown on the drawings and specified. Unless otherwise noted 2 pole, 3 wire, rated at 20 amps, 120 Volt, Hospital Grade shall be provided. All receptacles include safety plates.

- **Duplex receptacles** shall be NEMA style 5-20R, color Ivory for use on normal power circuits, and color Red for use on critical branch power circuits.
 - **Simplex receptacles** shall be NEMA style 5-20R, color Ivory for use on normal power circuits, and color Red for use on critical branch power circuits.
 - **Safety receptacles** (if required) shall be duplex type, be NEMA style 5-20R, color Ivory for use on normal power circuits, and color Red for use on critical branch power circuits.
- Receptacles shall limit proper access to energized contacts.

GROUNDING JACKS

Externally accessible grounding jacks shall be provided in the Security headwall wall unit as shown in these shop drawings. The solid brass receptacle shall be enclosed in a non-conductive housing and shall be spring loaded, with a twist-to-lock action and shall be manufactured to the requirements of NEC article 517, and NFPA 99.

SWITCHING

Switches shall be Industrial Grade 120/227 volt, 20 amps. Switch type options include SPST, 3-Way or Momentary, and shall be provided as shown on submittal shop drawings. Low Voltage Switching will be 0-12 volts, 15 amps unless otherwise noted. HSI shall furnish, pre-install and wire all switches.

REACT™ CLOCK TIMER

When indicated on the drawings, a HSI React Clock Timer shall be provided and factory installed in the SECURITY HEADWALL unit. The circuitry shall allow the activation of the elapsed time indicator by manually depressing the "Start" switch, or by a patient ventricular alarm condition broadcast through the bedside physiological monitor. Manually operated "Stop/Start", "Reset", and "Mode" switches are also provided and pre-wired within the React unit. The installing contractor shall perform wiring and electrical actuation between the monitor and the React unit.

B. Med Gas

MEDICAL GAS PIPING

Oxygen, vacuum, medical air, evacuation, and nitrous oxide outlets shall be as indicated in plans and specifications. All outlets will be installed by HSI at the factory, manifolded and tested in accordance with NFPA 99c. All Medical Gas piping shall be Type L copper pipe. The mechanical contractor will bring the primary supply lines of the medical gases to the singular terminal connections at locations shown on the drawings.

100% USA Manufactured

Union Made

12 Year Product Warranty

All specifications are nominal and subject to change without notice. HSI makes no warranty, expressed or implied, except for the express written warranty extended on the sale of its products.

3. Components Cont.

C. Provisions

NURSE CALL EQUIPMENT

All nurse call equipment shall be furnished and installed by others as indicated in the plans and specifications. A labeled pull-cord shall be provided by HSI for use by the nurse call installer.

MONITORING EQUIPMENT

Monitor equipment shall be provided and installed by others. HSI shall provide a matching blank plate for the monitor receptacle to the monitor supplier when requested. The SECURITY HEADWALL unit will include a factory installed raceway and labeled pull cord to accommodate the monitor wiring. Where specified a monitor slide and optional pivot/tilt arm shall be provided to support the physiological monitor.

4. Installation

HSI shall supply drawings with locations of stub ups. The installer or contractor shall install and make primary gas and electrical connections. Installer or contractor shall also install steel security enclosure. Hardware to mount the headwall shall be provided by the installing contractor to meet local requirements. Instructions shall be furnished by HSI. Customer shall coordinate with Hospital Systems, Inc for equipment supplied by others to ensure compatibility of this equipment with the headwall. It shall be the responsibility of the mechanical contractor to perform and certify all pressure tests as required by NFPA 99c, and to install the gas outlet finishing assemblies. Please refer to the installation instructions for further details.

5. Ordering Information

All HSI headwalls are Build-to-Order and will be assigned their own unique model numbers during the submittal phase.

Please visit our website at www.HSIheadwalls.com for the contact information of your local sales representative.

100% USA Manufactured

Union Made

12 Year Product Warranty

All specifications are nominal and subject to change without notice. HSI makes no warranty, expressed or implied, except for the express written warranty extended on the sale of its products.

HSI—750 Garcia Ave—Pittsburg, CA 94565
phone: 925.427.7800 fax: 925.427.0800
www.HSIheadwalls.com